

Certificate of Analysis

Analytical Testing Report · PT-141

SAMPLE IDENTIFICATION

Analyte †	PT-141	Date requested	—
Client	Pure North Peptides	Date received	4 Sep 2026
Vendor / brand †	purenorthpeptides.com	Date tested	9 Sep 2026
Assigned reference †	PNP-PT141-1002	Date reported	12 Sep 2026
Molecular formula	C50H68N14O10	Sample ID	SAMPLE-PT141-091226
Theoretical mass	1025.16 g/mol	Quantity received †	One sealed vial, lyophilized
Condition on receipt	Intact and sealed on arrival, label legible.		

† Client-declared information. Results apply to the sample tested. Full reporting statements accompany the online verification record.

HPLC-UV PURITY & CONTENT ANALYSIS

METHOD · RP-HPLC-UV 220 NM · SPEC ≥ 98.0%

Expected content † LABEL CLAIM	10.00 mg
Mass SAMPLE AS RECEIVED	10.73 mg
Purity HPLC AREA PURITY	99.18%
Fill accuracy MASS OVER LABEL CLAIM	107.3%

PURITY CONFORMS

Client-supplied specification ≥ 98.0% for area purity: this sample meets that specification. Simple acceptance; no guard band. Applies to this specification only.



Fig. 1 · Product photograph

APPEARANCE

Appearance	White lyophilized powder
Cap colour	Green

DETECTION
220 nm
FLOW RATE
1.0 mL/min

COLUMN
C18, 4.6 x 150 mm, 5 µm
INJECTION
20 µL

MOBILE PHASE
Water / acetonitrile, 0.1% TFA
RUN TIME
25 min

GRADIENT
5 to 65% B over 20 min
INTEGRATION
Area % of total response

Content calibrated against PT-141 working standard, qualified in house. Calculation basis: As received, uncorrected for water and counter-ion content.

AUTHORISATION

CONTROLLED CERTIFICATE RECORD

Canada Peptide Testing
RECORD ENTRY

Canada Peptide Testing
CERTIFICATE RELEASE
2026-09-15

Verify this certificate

Scan for the current certificate, signed results and reporting statements.

canadapeptidetesting.com/certificates/CPT-PT141-10-091226

Record 08c0d90a5b899843118989c6 · signing key k1

Verification key 22SACQAP62BM



SCAN TO VERIFY